

## **EUTHANEX INSTRUCTION** (Blue Line)

- 1. Ensure green flowmeter for individual cages is turned off.**
- 2. Remove cage lids (leave wiretop in place).**
- 3. Place cages in Euthanex chamber & close.**
- 4. Select Adults (15 minute cycle) or Neonates (60 minute cycle).**
- 5. Press RUN.**
- 6. Do not open door until cycle is finished.**
- 7. Press STOP to abort cycle if necessary.**