with Oxivir spra/wipesbefore & after use. O₂ gaugeis above 100 PSI

Induction at 2-3% isoflurane vaporizer setting.	
Maintenanceat 1-2%isoflurane vaporizer setting.	
Oxygen flow ratesat 0.5 L/minutefor each chamber or Bains circuit in use.	
O₂ Flow Rate(L/min)	
0.5	
1.0	
1.5	
2.0	
2.5	
3.0	

SC #288 Eff. 9/23